

Blood Pressure Monitor Performance Report

Customer: KU-CTSU Rainbow		Contact: Andra Lahner														
Street: 3901 Rainbow Blvd (2nd Floor)		Phone # 913-574-3018														
City/ST/Zip: Kansas City, KS 66160		P= Pass F=Fail														
MFG	Model#	S/N	Electrical Safety Check	Condition Controls/ Switches	Condition Power Cord & Strin Relief	Hardware Integrity	Indecators & Display Operation	MFG. Labeling Intact	Normal		Hypertonic		Hypotonic		SPO2	Date Performed
									120/80	P 60	200/150	P130	60/30	P40	97%	

Blood Pressure Monitor Performance Report

Customer: KU-CTSU Rainbow		Contact: Andra Lahner														
Street: 3901 Rainbow Blvd (2nd Floor)		Phone #: 913-574-3018														
City/ST/Zip: Kansas City, KS 66160		P= Pass F=Fail														
MFG	Model#	S/N	Electrical Safety Check	Condition Controls/ Switches	Condition Power Cord & Strin Relief	Hardware Integrity	Indecators & Display Operation	MFG. Labeling Intact	Normal		Hypertonic		Hypotonic		SPO2	Date Performed
									120/80	P 60	200/150	P130	60/30	P40	97%	