

**Scale Performance Report**

Customer:		KU-CTSU Fairway				Contact:		Shelby Willams							
Street:		4350 Shawnee Mission Parkway 3rd Floor				Phone #		913-574-3018							
City/ST/Zip:		Fairway, KS 66205						P= Pass F=Fail							
MFG	Model#	S/N	Electrical Safety Check	Condition Controls/ Switches	Condition Power Cord & Strin Relief	Hardware Integrity	Indecators & Display Operation	MFG. Labeling Intact	Units of Measure: lbs						Date Performed
									Target	Actual	Target	Actual	Target	Actual	
Health-O-Meter	500KL	5000087608	P	P	P	P	P	P	50.0	50.0	250.0	250.0	500.0	500.0	01/31/2022
Health-O-Meter	500KL	5000135815	P	P	P	P	P	P	50.0	50.0	250.0	250.0	500.0	499.6	01/31/2022
Health-O-Meter	500KL	5000133809	P	P	P	P	P	P	50.0	50.0	250.0	250.0	500.0	500.2	01/31/2022
Detecto	MV1	E14019-0058	P	P	P	P	P	P	50.0	49.8	250.0	250.2	500.0	500.0	01/31/2022

Technicians Initials	JC		Technician:	Jim Cowart		Test equipment used	50lb Weights: Kansas Metrology Laboratory Report Number: 15825 Expires: 02/2023	
	Notes:			Next Due for Calibration 01/2023				

**Calibration Performed by Cobb Refrigeration and Lab Service 8071 Flint St. Lenexa KS, 66212 913-403-0088**

