

Test Results

Test Description	True Value	Test Result	Lower Limit	Upper Limit	Units		TUR
23.0 °C	23.00	23.1	22.7	23.3	°C	Pass	1.95
-190.0 °C	-190.00	-189.8	-190.7	-189.3	°C	Pass	2.06
990.0 °C	990.00	989.9	989.2	990.8	°C	Pass	3.06
1200 °C	1200.0	1200	1199	1201	°C	Pass	2.50

\*\*\*\*\*END OF CERTIFICATE\*\*\*\*\*



Signed:

*Bart A. Schwartz*  
 Bart A. Schwartz , Engineer in Charge

Port of Certification for

SERIAL NUMBER: 45470242WS

ASSET NUMBER: 45470242WS

1 On: Monday, July 12, 2021

Calibration Services Since 1962 Page 2 of 2

**Test Results** indicate the following: Found-Left: Unit was left as found. As-Left: Unit was left after adjustments.