

Scale Performance Report

Customer: KU-CTSU Rainbow			Contact: Andra Lahner														
Street: 3901 Rainbow Blvd (2nd Floor)			Phone #: 913-574-3018														
City/ST/Zip: Kansas City, KS 66160			P= Pass F=Fail														
MFG	Model#	S/N	Electrical Safety Check	Condition Controls/ Switches	Condition Power Cord & Strin Relief	Hardware Integrity	Indecators & Display Operation	MFG. Labeling Intact							Date Performed		
									Target	Actual	Target	Actual	Target	Actual			
Health-O-Meter	500KL	5000163137	P	P	P	P	P	P	50.0	50.2	250.0	251.0	500.0	501.5	01/19/2021		
Detecto	MV1	E17619-0049	P	P	P	P	P	P	50.0	50.2	250.0	251.8	500.0	500.2	01/19/2021		
Detecto	500KL	5000186170	P	P	P	P	P	P	50.0	50.1	250.0	250.1	500.0	500.2	01/19/2021		
DELP G040																	
Health-O-Meter	500KL	5000178353	P	P	P	P	P	P	50.0	50.1	250.0	250.2	500.0	500.0	01/19/2021		
Health-O-Meter	500KL	5000163141	P	P	P	P	P	P	50.0	50.1	250.0	250.2	500.0	500.1	01/19/2021		

Technicians
Initials

JC

Technician: Jim Cowart

Test equipment used
50lb Weights: Kansas Metrology Laboratory Report
Number: K16383 Expires: 12/29/2021
Reference Executive Order Per Expiration Date

Notes: Next Due for Calibration 01/2022

Calibration Performed by Cobb Refrigeration and Lab Service 8071 Flint St. Lenexa KS, 66212 913-403-0088

