

CERTIFICATION OF ACCURACY

Cobb Refrigeration & Laboratory Service

8071 Flint St, Overland Park, KS 66214

Tel: (913) 403-0088, Fax: (913) 403-0089

Customer Information:

Company Name: University of Kansas – CTSU Rainbow

Company Address: 3901 Rainbow, Kansas City KS 66160

Contact Person: Jodi Potts / Shelby Williams

Thermometer Information: Built in

Manufacture: Thermo Fisher Scientific

Model: TSX60086D

Serial No: 1124092101210920

Department/Location: CTSU / Delp G040

Temperature Range(°C): -80°C

Asset#: _____

Certified Temperature (°C) -78.4°C

Certified %CO2 N/A Certified RH% N/A

Observed Temperature (°C) -78°C

Observed %CO2: N/A Observed RH% N/A

Observer Accuracy (°C) <2°C

Passed/Failed: Pass

Certification Date: 10-4-2021

Certification Due: 10/2022

Note: *Acceptable temperature range will be determined by the product criteria stored in unit.

Service Technician: John Greer

Date: 10-4-2021

Certifying Equipment: Fluke 52II, Serial: 21960025, Due date: 5-24-2022

Review By: _____ Date: _____ Manager: _____ Date: _____